

Test certificate

Dato	<u>03.07.2009</u>
Test operator	<u>Vidar</u>
Serial number	<u>0906007</u>

Software rev	<u>2.0.3.1</u>
Firmware rev	<u>2.0</u>
Hardware rev	<u>5</u>

Test

Calibration of pressure sensor	<u>Ok</u>
Test of shunt interface	<u>Ok</u>
Maximum error @ 2,5b [3-21 l/min]	<u>0,5</u>
Maximum error @ 4b [3-27 l/min]	<u>0,9</u>
Maximum error @ 5b [3-27 l/min]	<u>0,6</u>
Calibration	<u>Ok</u>

Measured accuracy

